



Intended use

## Intended use as medical device

For ovens covered by the scope of the Directive 93/42/EEC (Directive of the Commission on the harmonization of the legal regulations of the Member States on medical devices) the following intended purpose applies

Type	Intended use as medical device	Classification according to MDD	Option A6
SNplus, SFplus	The appliance is intended for sterilizing medical material through dry heat at atmospheric pressure.	IIb	-
SN, SF *)	The appliance is intended for sterilizing medical material through dry heat at atmospheric pressure.	IIb	yes
ICOmed	The appliance is intended for the creation and maintenance of constant environmental conditions for application in the field of in vitro fertilisation (IVF), especially for the incubation of oocytes, spermatozoa and zygotes in special culture dishes for IVF application as well as for gene expression and the biosynthesis of RNA and proteins.	IIa	-
UNmplus, UFmplus	The appliance may be used for heating fango, silicate and APS packs for physical therapy and keeping them warm.	I	-
UNm, INm, UFm, IFm *)	The appliance may be used for heating fango, silicate and APS packs for physical therapy and keeping them warm.	I	yes
INmplus, IFmplus	The appliance may be used for temperature control of rinsing and infusion solutions and contrast agents.	I	-
IFbw	The appliance may be used for heating non-sterile covers and cloths.	I	-

\*) Note: in order to meet all requirements of guideline 93/42/EEC attach. 1, an additional over temperature protection (option A6) is compulsory